

TERMOPRODUKT

CERTIFICATE OF CONFORMITY

Device model: Termioplus data logger
Range: -30.0 to 70.0°C
Range: 0% to 99%
Resolution: 0.1°C; 0.1%RH
Accuracy: $\pm 0.4^{\circ}\text{C}$: -10.0°C to 40.00°C Otherwise $\pm 1.0^{\circ}\text{C}$
Accuracy: $\pm 2\% \text{RH}$: 30.0% to 70.0% Otherwise $\pm 5\%$
Made in Poland


Conforms to RoHS & REACH Legislation
Conforms to EC Reg 1935/2004 & EU Reg 10/2011

DECLARATION OF CONFORMITY

We hereby declare that the following product: Termioplus is manufactured, tested, inspected and stored, to conform in all respects to the stated specifications and in accordance with the Quality System of Termoprodukt.

We also confirm that when used in an appropriate manner, the instruments are suitable for use within the Food Industry, and also, where appropriate and required, comply with EEC Directives with regard to European Legislation and CE Marking requirements. Including EC Regulation No 1935/2004 of 27 October 2004 on Materials and articles intended to come into contact with food; & Commission Regulation (EU) No 10/2011 on plastic materials and articles intended to come into contact with food. All Termoprodukt supplied instruments are suitable for use within HACCP plans where the quoted specification for critical control points match the performance of the instrument.

Signed, on behalf of Termoprodukt



Piotr Wróblewski
Manager Owner

Valid as of date of purchase.