

## UK Declaration of Conformity


**Type:** Ranger Count 3000 RC31P... series scales  
**Model:** RC31P1502, RC31P3, RC31P6, RC31P15, RC31P30  
**Manufacturer:** OHAUS Corporation  
 8 Campus Drive #105, Parsippany, NJ 07054, United States of America

This declaration of conformity is issued under the sole responsibility of the manufacturer.

The object of the declaration described above is in conformity with the following applicable relevant statutory requirements:

Marking	UK Legislation	Standards / Norms	Certificates
	Electromagnetic Compatibility Regulations 2016	EN 61326-1:2013	
	Electrical Equipment (Safety) Regulations 2016	EN 61010-1:2010 +A1:2019	
	The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012	EN IEC 63000:2018	
<b>Weighing instruments for purposes according to Part 1, Article 3 (a to f) of The Non-automatic Weighing Instruments Regulations 2016</b>			
	Non-automatic Weighing Instruments Regulations 2016	EN 45501:2015	T8252-UK <sup>2)</sup>
1) XX - Year of affixing UKCA mark. Approved Body No. 8506 - NMI Certin UK (TIC) Ltd. Certification Body, conformity to type based on quality assurance of the production process (Module D), certificate UKCA-396 2) Approved Body No. 8506 - NMI Certin UK (TIC) Ltd., performed the evaluation of type examination and issued the certificate T8252-UK (Module B).			

Issued on: 2023-08-15



Chever Jean-Yves  
 President



Gavin Cao  
 Compliance Manager

