

Testo SE & Co KGaA, Celsiusstraße 2, 79822 Titisee-Neustadt, Germany

May 6, 2024

Declaration of conformity according to Reg. (EU) 1935/2004 for articles coming into contact with foods.

This is to declare that the following products:

0560 0110 testo 110 (Handheld Device)	
0516 0224 (Top Safe)	0615 1712 (Robust Air Probe)
0615 1212 (Penetration Temperature Probe)	0618 0072 (Robust Air Probe)
0618 0275 (Penetration Temperature Probe)	0572 2163 (Cable Temperature Probe)
0618 0073 (Penetration Temperature Probe)	0615 1912 (Waterproof surface probe)
0618 0071 (Flexible Temperature Probe)	0615 2211 (Stainless steel food probe (IP65))
0615 2411 (Robust food penetration probe)	0615 3211 (Frozen food probe for screw-in use)
0615 3311 (Stainless steel food probe (IP67))	

conform to the regulation of the Directive (EU) 1935/2004 of the European Parliament from the 27th of October 2004 and the requirements of the LFGB (German Food and Feed Code) §30, §31 in the current version.

Good manufacturing practice for the materials and items we use that are intended to come into contact with food complies with Regulation (EU) No. 2023/2006 (GMP) of December 22, 2006.

Description:

Food contact materials are all articles which are intended to come directly or indirectly into contact with foodstuffs. In measuring instruments, these are only the parts whose intended purpose, specified by the manufacturer, is to have permanent contact with food.

The requirements of the Regulation Reg. EU 1935/2004 thus refer to the measurement tip of plug-in and fixable probes in all portable Testo temperature measuring instruments (see list above) which are immersed and/or penetrated into the food up to within 1 cm of the handgrip. For instruments with a fixed probe, this is the probe shaft up to within 1 cm before the crossover to the plastic case. The exact description of the penetration depth and the intended purposes (in the food sector) can be found in the instruction manuals of the respective products.

INFORMATION FOR...

Temperature probes:

All temperature probes from Testo consist of food-quality stainless steel (steel key 1.4571, previously V4A) or comparable material. Probe handles, cables, connectors, housing, protective covers and similar consist of materials which are permitted to come into contact with food for a short period, but are not designed for long-term contact with food.

Other Testo measuring instruments:

Data loggers, infrared instruments as well as accessories such as protective covers, chargers, etc. are not articles for daily use falling within the sense of the regulation, as these are not designed to come into direct contact with foodstuffs.

Traceability:

Our products which can potentially come into contact with food are provided with a manufacturer part number and date of manufacture. This allows us to trace the production date as well as the delivery batches which is ensured by our ISO 9001 quality management system.

This document was created electronically and is valid without a signature.

Testo SE & Co. KGaA, HQ Titisee-Neustadt, Commercial Register Freiburg i. Br. HRB 715057; Supervisory Board Chairperson: Jürgen Hinn. General Partner: *Testo Management SE*, HQ Titisee-Neustadt, Commercial Register Freiburg i. Br. HRB 714984; Directors: Prof. Burkart Knospe (CEO), Eckhard Kloth, Peter Kräuter; Supervisory Board Chairperson: Jürgen Hinn.