



EA MLA Signatory
Český institut pro akreditaci, o.p.s.
Olšanská 54/3, 130 00 Praha 3

issues

according to section 16 of Act No. 22/1997 Coll., on technical requirements for products, as amended

CERTIFICATE OF ACCREDITATION

No. 48/2022

KALIST AKL s.r.o.
with registered office č.p. 8, 769 01 Třebětice, Company Registration No. 04432436

to the Calibration Laboratory No. 2394
KALIST AKL s.r.o., Calibration Laboratory

Scope of accreditation:

Calibration in the field of mass, volume, temperature, humidity, mechanical motion – rotational speed and time to the extent as specified in the appendix to this Certificate.

This Certificate of Accreditation is a proof of Accreditation issued on the basis of assessment of fulfillment of the accreditation criteria in accordance with

ČSN EN ISO/IEC 17025:2018

In its activities performed within the scope and for the period of validity of this Certificate, the Body is entitled to refer to this Certificate, provided that the accreditation is not suspended and the Body meets the specified accreditation requirements in accordance with the relevant regulations applicable to the activity of an accredited Conformity Assessment Body.

This Certificate of Accreditation replaces, to the full extent, Certificate No.: 580/2020 of 25. 9. 2020, or any administrative acts building upon it.

The Certificate of Accreditation is valid until: **25. 9. 2025**

Prague: 1. 2. 2022



Lukáš Burda
Director of the Department
of Testing and Calibration Laboratories
Czech Accreditation Institute
Public Service Company



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

KALIST AKL s.r.o.
KALIST AKL s.r.o., Calibration Laboratory
č.p. 8, 769 01 Třebětice

CMC for the field of measured quantity: Volume

| Ord. number 1 | Calibrated quantity / Subject of calibration | Nominal range | | Parameter(s) of the meas. quantity | Lowest expanded measurement uncertainty specified ^{2,4} | Calibration principle | Calibration procedure identification ³ | Work place |
|---------------------|---|---------------------|-------------------------------|--|--|-----------------------|--|---------------|
| | | min. unit | max. unit | | | | | |
| 1 | Piston pipettes and other piston volume meters | 0,5 µl 10,000 µl | to 10,000 µl to 100,000 µl | Distilled water | 0,13 % + 0,01 µl 0,05 % | Gravimetric method | KP-05 (ČSN EN ISO 8655-6, EURAMET cg-19) | |

- 1 Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- 2 The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.
- 3 If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).
- 4 The lowest uncertainty includes the effect of the operator and does not include the statistical components of uncertainty.




Accredited entity according to ČSN EN ISO/IEC 17025:2018:

KALIST AKL s.r.o.
KALIST AKL s.r.o., Calibration Laboratory
č.p. 8, 769 01 Třebětice

CMC for the field of measured quantity: Mass

| Ord. number 1 | Calibrated quantity / Subject of calibration | Nominal range | | Parameter(s) of the meas. quantity | Lowest expanded measurement uncertainty specified ^{2, 4} | Calibration principle | Calibration procedure identification 3 | Work-place |
|------------------|---|---|--|--|--|--|---|------------|
| | | min. unit | max. unit | | | | | |
| 1 * | Balances with non-automatic function | 1 mg 66 kg 270 kg | to to to | 66 kg 270 kg 11,00 0 kg | By E2 weight By F2 weight By M1 weight | 9.3 · 10 ⁻⁷ 9.3 · 10 ⁻⁶ 2.9 · 10 ⁻⁵ | Loading using a reference weight Loading using M1 reference weight and substitute load | KP-01 |
| 2 | Class F1, F2, M1, M2, M3 weights (according to OIML R111), reference weights, special weights and other bodies with constant mass | 1 mg 50 mg 1 g 5 g 50 g 200 g 2 kg 10 kg 20 kg 50 kg | to to to to to to to to to to | 50 mg 1 g 5 g 50 g 200 g 2 kg 10 kg 20 kg 50 kg 60 kg | 0.008 mg 0.016 mg 0.025 mg 0.06 mg 0.2 mg 2 mg 12 mg 25 mg 50 mg 100 mg | Comparison with a reference weight | KP-06 | |



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

KALIST AKL s.r.o.
KALIST AKL s.r.o., Calibration Laboratory
č.p. 8, 769 01 Třebětice

- 1 Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- 2 The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.
- 3 If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).
- 4 The lowest expanded measurement uncertainty is stated without accounting for the effect of the calibrated meter.



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

KALIST AKL s.r.o.
KALIST AKL s.r.o., Calibration Laboratory
č.p. 8, 769 01 Třebětice

CMC for the field of measured quantity: Mechanical motion

| Ord. number ¹ | Calibrated quantity / Subject of calibration | Nominal range | | | Parameter(s) of the meas. quantity | Lowest expanded measurement uncertainty specified ² | Calibration principle | Calibration procedure identification ³ | Work-place |
|--------------------------|--|--|------|--|------------------------------------|--|--|---|------------|
| | | min. | unit | max. | | | | | |
| 1* | Speed / rpm gauges | 10 min ⁻¹ 10,000 min ⁻¹ | | to 10,000 min ⁻¹ to 50,000 min ⁻¹ | | 2 min ⁻¹ 0.012 % + 1 min ⁻¹ | Comparison of an rpm gauge in a calibrated device with a reference rpm gauge | KP-02 | |

- 1 Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- 2 The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.
- 3 If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

KALIST AKL s.r.o.
KALIST AKL s.r.o., Calibration Laboratory
č.p. 8, 769 01 Třebětice

CMC for the field of measured quantity: Temperature

| Ord. num-ber ¹ | Calibrated quantity / Subject of calibration | Nominal range | | Parameter(s) of the meas. quantity | Lowest expanded measurement uncertainty specified ² | Calibration principle | Calibration procedure identification ³ | Work-place |
|---------------------------|--|--|----------------------------|--|--|---|---|------------|
| | | min. unit | max. unit | | | | | |
| 1 * | Glass thermometers | -40 °C 20 °C 150 °C | to to to | 20 °C 150 °C 180 °C | 0.15 °C 0.10 °C 0.12 °C | Comparison with a reference thermometer in a liquid bath | KP-03 part C | |
| | Indicating thermometers, temperature measuring chains, dataloggers | -196 °C | | | 0.70 °C | Comparison with a reference thermometer in liquid nitrogen | KP-03 part A | |
| | | -70 °C -40 °C 150 °C | to to to | -40 °C 150 °C 180 °C | 0.45 °C 0.40 °C 0.45 °C | Comparison with a reference thermometer in a climatic chamber | | |
| | | -40 °C 20 °C 150 °C | to to to | 20 °C 150 °C 180 °C | 0.15 °C 0.10 °C 0.12 °C | Comparison with a reference thermometer in a liquid bath | | |
| | | 180 °C 230 °C 415 °C 600 °C 800 °C | to to to to to | 230 °C 415 °C 600 °C 800 °C 1,100 °C | 0.25 °C 0.45 °C 1.5 °C 2.6 °C 2.8 °C | Comparison with a reference thermometer in a dry block | | |

Accredited entity according to ČSN EN ISO/IEC 17025:2018:

KALIST AKL s.r.o.
KALIST AKL s.r.o., Calibration Laboratory
č.p. 8, 769 01 Třebětice

| Ord. number ¹ | Calibrated quantity / Subject of calibration | Nominal range | | Parameter(s) of the meas. quantity | Lowest expanded measurement uncertainty specified ² | Calibration principle | Calibration procedure identification ³ | Work-place |
|--------------------------|--|---------------|-------------|------------------------------------|--|---|---|------------|
| | | min. unit | max. unit | | | | | |
| | Calibration of thermal equipment with temperature control | -196 °C | to -70 °C | | 0.90 °C | Comparison with a reference thermometer | KP-03 part B | |
| | | -70 °C | to -40 °C | | 0.60 °C | | | |
| | | -40 °C | to 150 °C | | 0.30 °C | | | |
| | | 150 °C | to 230 °C | | 0.45 °C | | | |
| | | 230 °C | to 415 °C | | 0.50 °C | | | |
| | | 415 °C | to 600 °C | | 1.6 °C | | | |
| | | 600 °C | to 800 °C | | 2.6 °C | | | |
| | | 800 °C | to 1,100 °C | | 2.8 °C | | | |
| | Temperature / Calibration of infrared non-contact thermometers | -30 °C | to 0 °C | | 2.2 °C | Comparison with a reference standard | KP-03-1R | |
| | | 0 °C | to 20 °C | | 1.5 °C | | | |
| | | 20 °C | to 80 °C | | 1.2 °C | | | |
| | | 80 °C | to 200 °C | | 1.6 °C | | | |
| | | 200 °C | to 350 °C | | 2.5 °C | | | |
| | | 350 °C | to 500 °C | | 3.0 °C | | | |

¹ Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.

² The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.

³ If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

KALIST AKL s.r.o.
KALIST AKL s.r.o., Calibration Laboratory
č.p. 8, 769 01 Třebětice

CMC for the field of measured quantity: Time and frequency quantities

| Ord. number ¹ | Calibrated quantity / Subject of calibration | Nominal range | | Parameter(s) of the meas. quantity | Lowest expanded measurement uncertainty specified ² | Calibration principle | Calibration procedure identification ³ | Work-place |
|--------------------------|--|---------------|-------------|------------------------------------|--|--|---|------------|
| | | min. unit | max. unit | | | | | |
| 1* | Time interval / stopwatches, timers and chronometers | 1 s | to 86,400 s | | 0.3 s | Comparison with reference stopwatches, manual activation | KP-07 | |

- 1 Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- 2 The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.
- 3 If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).



Accredited entity according to ČSN EN ISO/IEC 17025:2018:

KALIST AKL s.r.o.
KALIST AKL s.r.o., Calibration Laboratory
č.p. 8, 769 01 Třebětice

CMC for the field of measured quantity: Humidity

| Ord. number 1 | Calibrated quantity / Subject of calibration | Nominal range | | Parameter(s) of the meas. quantity | Lowest expanded measurement uncertainty specified ² | Calibration principle | Calibration procedure identification 3 | Work- place |
|---------------------|---|-------------------------------|--|--|--|--|---|----------------|
| | | min. unit | max. unit | | | | | |
| 1 * | Relative humidity / hygrometers and humidity measuring chains, humidity dataloggers | 10 % RH 65 % RH 80 % RH | to 65 % RH to 80 % RH to 95 % RH | | 1.8 % RH 1.9 % RH 2.2 % RH | Comparison with a reference hygrometer | KP-04 | |

- 1 Asterisk at the ordinal number identifies the calibrations, which the Laboratory is qualified to carry out outside the permanent laboratory premises.
- 2 The expanded measurement uncertainty is in accordance with ILAC-P14 and EA-4/02, part of CMC, and it is the lowest value of the respective uncertainty. If not stated otherwise, its coverage probability is approx. 95%. If not stated otherwise, the uncertainty values stated without a unit are relative to the value measured. If the calibration is carried out outside the laboratory premises, the measurement uncertainty may be affected.
- 3 If the document identifying the calibration procedure is dated, only these specific procedures are used. If the document identifying the calibration procedure is not dated, the latest edition of the specified procedure is used (including any changes).
RH – Relative Humidity

